

## PRODUCT DATA SHEET

Standing: 2022-04-27

LAUDA Varioshake VS 8 OE  
Shaker 230 V; 50/60 Hz  
Part Number: L003055

### Features

- Shaker with orbital motion, heavy duty, low noise, in a compact design
- Electronic speed control, stepless, gentle start-up
- Clear control panel for easy operation with rotary knobs
- Timer adjustable up to 60 min. or continuous operation
- Shaking table made of anodized aluminum with four plastic pins
- Drive by an overload-protected AC motor
- Compact, low-wear mechanism with mass balancing for a secure stand
- Constant speed, independent of load
- Excellent for continuous use
- Outer housing made of electrolytically galvanized sheet steel, powder-coated



### Technical Features

Ambient temperature range	10 ... 60 °C
Shaking tray (WxD)	330 x 330 mm
Power consumption max.	0.07 kW
Shaking amplitude	10 mm
Shaking frequency	20 ... 500 1/min
Load max.	8 kg
Overall dimensions (WxDxH)	350 x 375 x 160 mm
Weight	13 kg
Power plug	Power cord with angled plug (CEE7/7)
Power supply	230 V; 50/60 Hz



**Helago-CZ s.r.o.**  
Kladská 1082  
500 03 Hradec Králové  
Tel.: 495 220 229, 495 220 394  
Fax: 495 220 154  
<http://www.helago-cz.cz>  
E-mail: [info@helago-cz.cz](mailto:info@helago-cz.cz)

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0 • F + 49 (0) 9343 503-222  
[info@lauda.de](mailto:info@lauda.de) • [www.lauda.de](http://www.lauda.de)  
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Ralf Hermann, Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser